

***Cerebral oxygenation and blood flow in term infants during postnatal transition: BabyLux project.***

*This article has been accepted for publication in Disease in Childhood 2019 following peer review, and the Version of Record can be accessed online at <http://dx.doi.org/10.1136/archdischild-2018-316400>. ©*

*BMJ Publishing Group Ltd 2019*

# Archives of Disease in Childhood

## Cerebral oxygenation and blood flow in term infants during postnatal transition: BabyLux project.

Journal:	<i>Archives of Disease in Childhood</i>
Manuscript ID	archdischild-2018-316400.R1
Article Type:	Original article
Edition:	not in use
Date Submitted by the Author:	09-Feb-2019
Complete List of Authors:	De Carli, Agnese; Fondazione IRCCS Ca' Granda Ospedale Maggiore Policlinico, NICU Andresen, Björn; Rigshospitalet, Department of Neonatology Giovannella, Martina; The Barcelona Institute of Science and Technology, ICFO, Institut de Ciències Fotòniques Durduran, Turgut; The Barcelona Institute of Science and Technology, ICFO, Institut de Ciències Fotòniques ; Institució Catalana de Recerca i Estudis Avançats (ICREA) Contini, Davide; Politecnico di Milano, Dipartimento di Fisica Spinelli, Lorenzo; Istituto di Fotonica e Nanotecnologie, Consiglio Nazionale delle Ricerche Weigel, Udo Michael; HemoPhotonics S.L. Passera, Sofia; Fondazione IRCCS Ca' Granda Ospedale Maggiore Policlinico, NICU Pesenti, Nicola; Fondazione IRCCS Ca' Granda Ospedale Maggiore Policlinico, NICU Mosca, Fabio; Fondazione IRCCS Ca' Granda Ospedale Maggiore Policlinico, NICU; University of Milan, Department of Clinical Sciences and Community Health Torricelli, Alessandro; Politecnico di Milano, Dipartimento di Fisica; Istituto di Fotonica e Nanotecnologie, Consiglio Nazionale delle Ricerche Fumagalli, Monica; Fondazione IRCCS Ca' Granda Ospedale Maggiore Policlinico, NICU; University of Milan, Department of Clinical Sciences and Community Health Greisen, Gorm; Rigshospitalet, Department of Neonatology
Keywords:	time-resolved spectroscopy, diffuse correlation spectroscopy, tissue oxygen saturation, cerebral blood flow, term infants

SCHOLARONE™  
Manuscripts

**TITLE:**

Cerebral oxygenation and blood flow in term infants during postnatal transition: BabyLux project.

Agnese De Carli<sup>1</sup>, Bjørn Andresen<sup>3</sup>, Martina Giovannella<sup>4</sup>, Turgut Durduran<sup>4,5</sup>, Davide Contini<sup>6</sup>, Lorenzo Spinelli<sup>7</sup>, Udo Michael Weigel<sup>8</sup>, Sofia Passera<sup>1</sup>, Nicola Pesenti<sup>1</sup>, Fabio Mosca<sup>1,2</sup>, Alessandro Torricelli<sup>6,7</sup> & Monica Fumagalli<sup>1,2\*</sup>, Gorm Greisen<sup>3</sup>.

1. *NICU Fondazione IRCCS Ca' Granda Ospedale Maggiore Policlinico, Milan, Italy.*
2. *University of Milan, Department of Clinical Sciences and Community Health, Milan, Italy.*
3. *Department of Neonatology, Rigshospitalet, Copenhagen, Denmark.*
4. *ICFO-Institut de Ciències Fotòniques, The Barcelona Institute of Science and Technology, Barcelona, Spain.*
5. *Institució Catalana de Recerca i Estudis Avançats (ICREA), Barcelona, Spain.*
6. *Politecnico di Milano-Dipartimento di Fisica, Milan, Italy.*
7. *Istituto di Fotonica e Nanotecnologie, Consiglio Nazionale delle Ricerche, Milan, Italy,*
8. *HemoPhotonics S.L., Castelldefels (Barcelona), Spain.*

*\*Corresponding author: Monica Fumagalli, MD; NICU Fondazione IRCCS Ca' Granda Ospedale Maggiore Policlinico, via della Commenda 12, 20122 Milan Italy – [monica.fumagalli@unimi.it](mailto:monica.fumagalli@unimi.it)*

**Word count: 3299**

The protocol is registered at ClinicalTrials.gov, identifier NCT02815618.

## ABSTRACT

### Objectives:

A new device that combines, for the first time, two photonic technologies (time-resolved near-infrared spectroscopy and diffuse-correlation spectroscopy), was provided and tested within the BabyLux project. Aim was to validate the expected changes in cerebral oxygenation and blood flow.

### Methods:

A pulse oximeter and the BabyLux device were held in place (right hand/wrist and fronto-parietal region, respectively) for 10 minutes after birth in healthy term infants delivered by elective cesarean section. Pulse oximeter saturation( $SpO_2$ ), cerebral tissue oxygen saturation( $StO_2$ ) and blood flow index(BFI) were measured over time. Tissue oxygen extraction(TOE) and cerebral metabolic rate of oxygen index(CMRO<sub>2</sub>I) were calculated.

### Results:

Thirty infants were enrolled in two centers. After validity check of data, 23% of infants were excluded from TOE and CMRO<sub>2</sub>I calculation due to missing data. As expected,  $SpO_2$  (estimate 3.05 %/min; CI: 2.78, 3.31 %/min) and  $StO_2$  (estimate 3.95 %/min; CI: 3.63, 4.27 %/min) increased in the first 10 min after birth, whereas BFI (estimate  $-2.84 \cdot 10^{-9}$  cm<sup>2</sup>/s/min; CI:  $-2.50 \cdot 10^{-9}$ ,  $-3.24 \cdot 10^{-9}$  cm<sup>2</sup>/s/min) and TOE (estimate -0.78%/min; CI: -1.12, -0.45 %/min) decreased. Surprisingly, CMRO<sub>2</sub>I decreased (estimate  $-7.94 \cdot 10^{-8}$ /min; CI:  $-6.26 \cdot 10^{-8}$ ,  $-9.62 \cdot 10^{-8}$ /min).

### Conclusions:

Brain oxygenation and BFI during transition were successfully and simultaneously obtained by the BabyLux device; no adverse effects were recorded and the BabyLux device didn't limit the standard care.

The preliminary results from clinical application of the BabyLux device are encouraging in terms of safety and feasibility; they are consistent with previous reports on brain oxygenation during transition although the interpretation of the decreasing CMRO<sub>2</sub>I remains open.

**KEY WORDS:**

Near-infrared spectroscopy, diffuse correlation spectroscopy, tissue oxygen saturation, cerebral blood flow, term infant.

**ABBREVIATIONS:**

Absorption coefficient ( $\mu_a$ ), arterial oxygen saturation ( $S_aO_2$ ), venous oxygen saturation ( $S_vO_2$ ), blood flow index (BFI), cerebral metabolic rate of oxygen index ( $CMRO_2I$ ), concentration of oxy-haemoglobin ( $O_2Hb$ ), concentration of deoxy-haemoglobin (HHb), diffuse correlation spectroscopy (DCS), distribution of time-of-flight (DTOF), near-infrared spectroscopy (NIRS), tissue oxygen extraction (TOE), pulse oxygen saturation ( $SpO_2$ ), reduced scattering coefficient ( $\mu_s'$ ), time-resolved spectroscopy (TRS), tissue oxygen saturation ( $StO_2$ ).

## INTRODUCTION:

During the last decades, progress in neonatal medicine has led to an increased survival rate of preterm infants. Despite this, the risk of brain damage and later neurodevelopmental deficits is still high and the understanding of the underlying pathophysiological mechanisms is nevertheless incomplete.[1,2]

Different perinatal factors (hemodynamics, oxygen metabolism, infection-inflammation) are involved in the pathogenesis of brain damage. A more accurate identification of the leading mechanisms on a single-patient basis is desirable to provide individualized care and targeted intervention aimed to safeguard the developing brain.

The most vulnerable period is represented by the first hours and days after birth due to the risk of haemodynamic disturbances occurring during the transitional circulation combined with the impact of respiratory distress syndrome. Furthermore, critically ill premature infants have impaired cerebral autoregulation, which may expose them to both hyperoxic and hypoxic insults, both involved in brain injury.[3,4]

To that end, a continuous and non-invasive monitoring of cerebral perfusion and oxygenation has been searched for.[5]

Commercially available near-infrared spectroscopy (NIRS) devices are currently used in clinical care and in clinical trials as cerebral oximeters.[6,7] Interpretation of cerebral oxygenation as a surrogate measure of cerebral blood flow, however, depends on the assumption of stable oxygen consumption. This is relevant for instance when using NIRS to estimate cerebral autoregulation capacity or when used clinically as an indication for interventions to increase blood flow. Alternatively cerebral oxygenation may be affected by change in oxygen demand. The BabyLux project aimed to provide a non-invasive and cot-side device, that combines time-resolved NIRS (TRS), for calculation of regional oxygenation, with newly developed diffuse correlation spectroscopy (DCS), for calculation of regional perfusion, in an attempt to resolve this ambiguity.

TRS measures the attenuation, delay, and the temporal broadening of relatively short light pulses (pulse duration ~100 ps) that has passed through a diffusive medium. TRS thus has the ability to

1  
2  
3 resolve path-lengths of photons and to separate the absorption and scattering coefficients allowing  
4  
5 for measurements of tissue oxygen saturation (StO<sub>2</sub>).[8]  
6

7 DCS studies the statistics of the diffuse coherent light into the tissue. The fluctuations of the intensity  
8  
9 measured at the surface of the tissue are affected by the movement of the moving scatterers, mainly  
10  
11 the red blood cells. Studying the statistics of these fluctuations allows for the quantification of the so-  
12  
13 called blood flow index BFI, a quantity proportional to the microvascular blood flow.[9,10]  
14

15 This innovative combined technology allows the calculation of cerebral metabolic rate of oxygen and,  
16  
17 to our knowledge, only few studies have been published on this topic using combined multi-distance  
18  
19 frequency domain NIRS and DCS.[11,12]  
20  
21

22  
23 As direct validation of cerebral oxygenation and blood flow against a ‘gold standard’ is not feasible  
24  
25 in newborn infants, we performed measurements in a clinical situation in which changes in cerebral  
26  
27 oxygenation and brain perfusion occur and have been previously described. The aim was to obtain  
28  
29 “expected results” according to the available evidence. We studied the postnatal transition of healthy  
30  
31 term infants delivered by elective, uncomplicated caesarian section; in this situation, cerebral tissue  
32  
33 oxygenation is expected to be lower than arterial saturation at all time-points, progressively rising  
34  
35 and reaching a plateau within the first 10 minutes after birth. Moreover, small-to-moderate changes  
36  
37 in cerebral blood flow, small changes in oxygen extraction, and no change in oxygen consumption  
38  
39 are expected given the observations of lower cord cortisol and catecholamines levels after elective  
40  
41 cesarean delivery compared to vaginal delivery. [13-15]  
42  
43  
44  
45

46 The BabyLux device provides a simultaneous measurement of regional oxygenation and perfusion  
47  
48 and allows to disentangle the interplay between oxygen demand and perfusion, resolving the  
49  
50 ambiguity in the interpretation of changes in regional saturation.  
51  
52

#### 53 **Aim:**

54  
55 The aim of this study was to validate the BabyLux device by simultaneously measuring the expected  
56  
57 changes in cerebral oxygenation and blood flow and by calculating the cerebral oxygen metabolism.  
58  
59

#### 60 **METHODS:**

**Study protocol:**

The study was conducted according to ISO 14155:2011 with external monitoring. Local research ethics committees approved the same study protocol in both centers (Rigshospitalet, Copenhagen, Denmark and Fondazione IRCCS Ca' Granda, Ospedale Maggiore Policlinico, Milan, Italy). The protocol is registered at ClinicalTrials.gov, identifier NCT02815618.

Inclusion criteria were: gestational age more than 37 weeks and planned delivery by uncomplicated elective cesarean section. Exclusion criteria were: congenital malformation apparent at birth with need for any additional assistance immediately following delivery and need for resuscitation or supplementary oxygen during the first 10 minutes after birth.

Parental consent was obtained before elective caesarean section.

The BabyLux device and the pulse oximeter (Radical 7; Masimo Corporation, Irvine, CA, U.S.A.) were both synchronized with the local clock time of the delivery room.

After birth the infant was immediately wrapped in warm towels and the head and right hand or wrist were cleaned to remove vernix and amniotic fluid (which could affect probe contact and signal quality). As soon as possible, the BabyLux probe was positioned in the fronto-parietal region of the newborn's head to measure  $StO_2$  and BFI, and the pulse oximeter probe was placed on the right hand or wrist to measure pulse oxygen saturation ( $SpO_2$ ) and pulse rate. Both probes were held in place by self-adhesive elastic bandage. For both instruments measurements lasted for at least 10 min, while standard neonatal care was given. According to international guidelines [16] we considered standard neonatal care: to warm and maintain normal temperature, to position the head in "sniffing" position, to clear secretion if needed, and to dry. If resuscitation was performed (from tactile stimulation onwards), according to our methods, the infant was excluded from the study.

In both centers timing of cord clamping was between 30 and 60 sec.

The following parameters were finally recorded over time by 10 s bins:  $SpO_2$ , pulse rate,  $StO_2$ , BFI.

Tissue oxygen extraction (TOE) was calculated as  $TOE = SpO_2 - StO_2$ , and the cerebral metabolic rate of oxygen index ( $CMRO_2I$ ) was calculated as  $CMRO_2I = TOE \cdot BFI$

1  
2  
3 Neonatal characteristics (gestational age, birth weight, sex, Apgar score) were collected.  
4

5 **Instrumentation:**  
6

7  
8 Two identical prototypes were built within the project and they were approved for research use  
9  
10 according to the clinical investigation plan by the national Medical Device Agencies in both countries.  
11  
12 The BabyLux device integrates TRS and DCS modules similar to those previously described by  
13  
14 Torricelli *et al.*[8] and Durduran *et al.*[9], respectively. In brief, the TRS module employs pulsed  
15  
16 lasers operating at three different wavelengths centered at about 685 nm, 760 nm, and 820 nm,  
17  
18 respectively. The pulse duration is <100 ps, with a repetition rate of 20 MHz, and an average output  
19  
20 power <1 mW for all wavelengths. A photomultiplier and a time-correlated single photon counting  
21  
22 board are used to acquire the distribution of photon time of flight (DTOF) for each wavelength. The  
23  
24 DCS module uses a continuous wave long coherence laser at 785 nm with an output power < 20 mW.  
25  
26 Two single avalanche photodiodes and a custom-made correlator unit are used to acquire the  
27  
28 autocorrelation of the measured light intensity. TRS and DCS share a compact and light weight fiber-  
29  
30 optic probe, for injection and collection of the light signals into the tissue. TRS and DCS source-  
31  
32 detector separation is 15 mm.  
33  
34  
35  
36

37 We have carefully evaluated the safety risk associated with the use of the pulsed laser (TRS) and  
38  
39 continuous wave laser (DCS) according to the standard IEC 60825-1:2007. Regarding the effect on  
40  
41 skin, for both lasers the emitting power is lower than the maximum permissible exposure (MPE) by  
42  
43 an order of magnitude. However, as an additional precaution for heat dissipation, the DCS laser is  
44  
45 switched off for 1 second every 9 seconds. For the effect on eye, the pulsed lasers are safe for the  
46  
47 operator because time exposure to the laser light is shorter (0.25 s) due to the blink reflex. Conversely,  
48  
49 for infants we can not rely on the blink reflex and/or ocular movement, therefore we cannot consider  
50  
51 the pulsed lasers eye safe for the accidental exposure to laser light. The same happens for the DCS  
52  
53 laser. In order to ensure safe operation of the device, proper measure of protection (i.e. a mask  
54  
55 covering the eyes like the mask used for phototherapy) was used together with a capacitive sensor  
56  
57  
58  
59  
60

1  
2  
3 designed to detect skin contact which is integrated in the probe head. These procedures were approved  
4  
5 by the ethical committees and by the national medical device agencies in Italy and Denmark.  
6

7 A detailed description of the BabyLux system is available in Giovannella *et al.*[17]  
8  
9

### 10 **Data processing and quality assessment:**

11  
12 Processing of TRS data consists of estimating the optical properties (absorption coefficient,  $\mu_a$ , and  
13  
14 reduced scattering coefficient,  $\mu_s'$ ) at all wavelengths, then calculating the hemodynamic parameters  
15  
16 (concentration of oxy-haemoglobin, O<sub>2</sub>Hb, and deoxy-haemoglobin, HHb), and finally evaluating the  
17  
18 quality of results.  
19

20  
21 For the estimation of the optical properties, the DTOF is fitted with a model for photon diffusion in  
22  
23 a semi-infinite homogeneous medium, after convolution with the instrument response function. The  
24  
25 fitting procedure is described in Cubeddu *et al.*[18]  
26

27  
28 Once the estimates of  $\mu_a$  at the three wavelengths are obtained, O<sub>2</sub>Hb and HHb are calculated by  
29  
30 means of Beer's law. Specific absorption values for hemoglobin are taken from the Prahl dataset.[19]  
31  
32 Lipid content in neonates is limited and therefore disregarded, while water content is fixed at  
33  
34 90%.[20]  
35

36  
37 From O<sub>2</sub>Hb and HHb values we obtained the total haemoglobin content  $tHb=O_2Hb+HHb$ , and  
38  
39  $StO_2=100 O_2Hb/tHb$ .  
40

41  
42 The main factor affecting the quality of TRS data is the number of photon in the DTOF(N). A  
43  
44 minimum value of  $N>10^3$  was used as threshold. The quality of fitting was evaluated through the  
45  
46 reduced chi-square  $\chi^2$  parameter. Large values ( $\chi^2>10$ ) and very low values ( $\chi^2<0.1$ ) were  
47  
48 discarded since they can be indicative of poor fitting model or signal-to-noise ratio, respectively.  
49  
50 Further criteria were set on the values for  $\mu_s'$  and on the values for StO<sub>2</sub>. Given the assumption of  
51  
52 photon diffusion (as a rule of thumb,  $\mu_s' \gg \mu_a$ ), if at any wavelength the fitting provides too low  
53  
54 values for  $\mu_s'$  (e.g.  $\mu_s' < 1 \text{ cm}^{-1}$ ), then there is the possibility that the use of the photon diffusion model  
55  
56 is inappropriate and values were therefore excluded.  
57  
58  
59  
60

1  
2  
3 The normalized intensity field autocorrelation curve acquired in DCS measurements is fitted to the  
4 solution of the diffusion equation for the autocorrelation function for the semi-infinite homogeneous  
5 geometry. As the source detector separation is known and the optical properties have been estimated  
6 by TRS at 760 nm, they can be inserted in the model, enabling determination of BFI. DCS curves  
7 acquired with an intensity rate below 10 kHz were excluded, due to poor signal-to-noise ratio. Results  
8 with residuals higher than 2 SD from the mean were rejected.  
9

10 For the calculation of BFI, 10 s moving average of  $\mu_a$  at 760 nm was used to reduce noise propagation  
11 from TRS to DCS analysis and a fixed sample average estimate of  $\mu_s'$  was used ( $\mu_s'=7 \text{ cm}^{-1}$ ); BFI  
12 and  $\mu_s'$  are indeed coupled in the equation describing the intensity field autocorrelation curve [21]  
13 and an eventual error in the latter is propagated in an error in the BFI [22]. Therefore using an  
14 individual estimation of  $\mu_s'$  can increase the inter-individual variability for the BFI.  
15

### 26 **Statistics:**

27 Descriptive analyses were obtained for the infants' neonatal characteristics. Continuous variables are  
28 described as mean (SD) or median (range), while categorical variables are expressed as number and  
29 percentage.  
30

31 For  $\text{SpO}_2$ ,  $\text{StO}_2$ , BFI, TOE and  $\text{CMRO}_2\text{I}$  box-plots were used to show changes over time.  
32

33 Over the first 10 min after birth the relationship between variables and time was studied using linear  
34 mixed-effect models with subject as random effect. Models results are expressed as estimate, 95% CI  
35 along with p-values. Values of  $p < 0.05$  were considered statistically significant. Logarithmic base 10  
36 transformation of BFI and  $\text{CMRO}_2\text{I}$  data was done to normalize the right-skewed distribution of  
37 residuals.  
38

39 From 10 min after birth onward, median value, 25<sup>th</sup> and 75<sup>th</sup> percentiles were calculated for each  
40 variable.  
41

42 Statistical analyses were performed using R, version 3.4.3 (R Foundation for Statistical Computing,  
43 Vienna, Austria).  
44  
45  
46  
47  
48  
49  
50  
51  
52  
53  
54  
55  
56  
57  
58  
59  
60

**RESULTS:**

Thirty infants were enrolled from October 2016 to April 2017: 14 infants in Copenhagen (CPH) and 16 in Milan (MI). Four infants were excluded: one requiring resuscitation at birth, two due to parental consent withdrawn, and one because of technical reasons (BabyLux software crash).

Mean (SD) gestational age was 38.4 (0.7) weeks [38.5 (0.8) in CPH and 38.3 (0.6) in MI]; mean (SD) birthweight was 3258g (393g) [3299g (438g) in CPH and 3225g (366g) in MI]; male (%) were 13 (50) [5 (45) in CPH and 8 (53) in MI]; median (range) Apgar 1' was 10 (8-10) [10 (9-10) in CPH and 9 (8-9) in MI]; median (range) Apgar 5' was 10 (9-10) [10 (10) in CPH and 9 (8-9) in MI].

After validity check of data, the final number of accepted measurements were 23 (88.5%) for SpO<sub>2</sub>/pulse rate (10 in CPH and 13 in MI); 23 (88.5%) for TRS (8 in CPH and 15 in MI), 25 (96.2%) for DCS (11 in CPH and 14 in MI), 20 (76.9%) for calculated TOE and CMRO<sub>2</sub>I.

Figures 1 to 5 show changes of the measured variables from the 3<sup>rd</sup> min after birth over time.

Data from birth to the 3<sup>rd</sup> min are not displayed as very few data-points were collected in that time-frame: the average starting time for BabyLux measurement was 3.5 min (SD 1.5 min) after birth (see details in supplementary material).

We analyzed changes in all parameters until the 10<sup>th</sup> min after birth: SpO<sub>2</sub> significantly increased over time, as well as StO<sub>2</sub> while BFI and TOE significantly decreased. Surprisingly, a significant decrease in CMRO<sub>2</sub>I was observed.

SpO<sub>2</sub> (estimate 3.05 %/min; CI: 2.78, 3.31 %/min; p<0.001) and StO<sub>2</sub> (estimate 3.95 %/min; CI: 3.63, 4.27 %/min; p<0.001) increased in the first 10 min after birth, whereas BFI (estimate  $-2.84 \cdot 10^{-9}$  cm<sup>2</sup>/s/min; CI:  $-2.50 \cdot 10^{-9}$ ,  $-3.24 \cdot 10^{-9}$  cm<sup>2</sup>/s/min; p<0.001) and TOE (estimate  $-0.78\%$ /min; CI:  $-1.12$ ,  $-0.45 \%$ /min; p<0.001) decreased. Also CMRO<sub>2</sub>I decreased (estimate  $-7.94 \cdot 10^{-8}$ /min; CI:  $-6.26 \cdot 10^{-8}$ ,  $-9.62 \cdot 10^{-8}$  /min; p<0.001).

From the 10<sup>th</sup> min onward, median SpO<sub>2</sub> (25<sup>th</sup> -75<sup>th</sup> percentiles) was 94.8% (92.9%-97.1%), StO<sub>2</sub> 65.6% (59.9%-73.3%), BFI  $2.50 \cdot 10^{-8}$  cm<sup>2</sup>/sec ( $1.88 \cdot 10^{-8}$ - $3.46 \cdot 10^{-8}$ ), TOE 29.0% (22.7%-31.4%) and CMRO<sub>2</sub>I  $6.09 \cdot 10^{-7}$  ( $4.78 \cdot 10^{-7}$ - $8.67 \cdot 10^{-7}$ ).

1  
2  
3 No adverse effects were recorded during the study period.  
4

## 5 **DISCUSSION:**

6  
7 The use of the BabyLux device during the transition period after birth was feasible; it was safe for  
8 the patients and did not limit the standard care.  
9

10 The recording start time for the BabyLux device was reasonable: we obtained valid measurements in  
11 58% of infants by 3 min after birth and in 88% of infants by 5 min. These results are similar to those  
12 reported by previous studies during transition: Urlesberger *et al.* measured 50% of infants by 3 min  
13 [23] using INVOS5100 and Baik *et al.* 79% of newborns by 3 min with NIRO.[24] Shorter starting  
14 time are reported by Ziehenberger E. *et al.*, with NIRO or INVOS5100C (mean, 95 sec. and 93 sec,  
15 respectively) [25], Fauchere *et al.* with NIRO (median, 2 min; range, 1-4 min),[26] Almaazmi *et al.*  
16 with FORESIGHT (median time to signal, 63 s; interquartile range, 38-88 s),[27] and Isobe with  
17 IMUC7000 (measurement begun at 1-2.5 min).[28]  
18  
19

20 After validity check of data, 23% of infants were excluded from TOE and CMRO<sub>2</sub> calculation due to  
21 missing data (TRS in 12% and DCS in 4% of case) suggesting that technical issues still need to be  
22 improved. However, other instruments, as pulse-oximeter, routinely used in clinical practice, showed  
23 similar performances.  
24  
25

26 The recorded SpO<sub>2</sub> data are in line with the reference values by Dawson *et al.* for infants born by  
27 elective cesarean section.[31]  
28  
29

30 The measured changes in StO<sub>2</sub> and BFI are consistent with the previous reports.[23-30,32]

31 Cerebral StO<sub>2</sub> by fronto-parietal TRS sensor followed the rise in SpO<sub>2</sub>, similarly to previous  
32 observations performed on the same population using other NIRS techniques.[23-30] Indeed, StO<sub>2</sub>  
33 values derived from TRS seem grossly comparable with values obtained by NIRO300,[24-26] FORE-  
34 SIGHT cerebral oximeter,[27] IMUC7000 [28] (although some of these studies censored data points  
35 where StO<sub>2</sub> exceeded SpO<sub>2</sub>, which we did not do) but lower than values measured by INVOS5100  
36 with neonatal sensors. [23,25,29,30] This observation is consistent with previous studies reporting  
37  
38  
39  
40  
41  
42  
43  
44  
45  
46  
47  
48  
49  
50  
51  
52  
53  
54  
55  
56  
57  
58  
59  
60

1  
2  
3 significantly higher regional StO<sub>2</sub> values when using pediatric and neonatal sensors compared to the  
4  
5 INVOS5100 adult sensors.[33,34]  
6

7  
8 The derived value of TOE decreased over time in the first 10 min after birth, although a high inter-  
9  
10 individual variability was observed. This finding is in line with previous studies that investigated  
11  
12 changes in fractional tissue oxygen extraction (FTOE) showing a reduction in the first minutes of life,  
13  
14 although TOE and FTOE cannot be directly compared during the sudden increase in oxygen delivery  
15  
16 after birth.[23,35]  
17

18  
19 Cerebral blood flow (expressed as BFI) decreased by 30% over the study time. This result is  
20  
21 consistent with a 30% decrease in cerebral blood flow velocity, measured by Doppler, previously  
22  
23 reported in healthy term newborns.[32] This data, however, was obtained after vaginal birth and the  
24  
25 decrease was observed from 7 to 13 min after birth. Nevertheless, a cerebrovascular response to the  
26  
27 increase in blood oxygen content at birth is to be expected.  
28  
29

30  
31 This is the first study in which DCS was used to measure BFI during transition after birth. BFI is an  
32  
33 absolute quantification of cerebral blood flow with an unusual unit (cm<sup>2</sup>/s) that differs from the  
34  
35 conventionally reported ml/100g/min for blood flow and cm/s for blood flow velocity. However, DCS  
36  
37 has been broadly qualitatively validated and a few in-vivo calibrations have been reported.[10] Using  
38  
39 a calibration obtained in young piglets by comparing the BabyLux DCS with CBF measured with  
40  
41 positron emission tomography,[38] we calculated that the median BFI value from the 10<sup>th</sup> minutes  
42  
43 after birth onward corresponded to 23.3 ml/100g/min. These results are in line with those expected  
44  
45 in healthy term infants, taking into consideration the reported coefficient of variability and the small  
46  
47 number of observations.[39,40]  
48  
49

50  
51 Furthermore, BFI values measured by the BabyLux device after the 10<sup>th</sup> min after birth are  
52  
53 comparable with previously reported BFI values obtained using a combination of frequency domain  
54  
55 NIRS and DCS systems in healthy newborns admitted to the nursery.[36,37] The BabyLux  
56  
57 technology offers a unique possibility to obtain direct measurements of cerebral blood flow even in  
58  
59 such a peculiar situation when no other available technology (PET, SPECT, MRI) could be used.  
60

1  
2  
3 We expected constant  $CMRO_2I$  values because infants born by uncomplicated cesarean section have  
4 low levels of catecholamine compared to those vaginally delivered and therefore are not expected to  
5 be significantly stressed or asphyxiated and therefore not to have developed an oxygen 'debt' during  
6 delivery.[14,15]  
7  
8  
9

10  
11  
12 Therefore, it was surprising that we found a nearly 50% decrease in  $CMRO_2I$ , that was highly  
13 statistically significant, from the first minutes after birth until stabilization at about 10 minutes.  
14

15  
16  $CMRO_2I$  is a measure of oxygen metabolism and thus reflects the infants' cerebral metabolic state.  
17 Arousal occurring at birth should be associated with higher  $CMRO_2I$ . Our results therefore may  
18 suggest that even an uncomplicated cesarean section involve some physical stress and strain and, due  
19 to the delay in the recording start time in our study, we might have measured the descending curve  
20 associated with the gradual reduction in the alert state with cerebral activation a few minutes after  
21 birth. Alternatively, we may speculate that this unexpected result is due to an oxygenation-level  
22 dependent error of measurement of at least one of involved variables  $SpO_2$ ,  $StO_2$  or BFI, and/or an  
23 oxygenation-level dependent change in the arterio-venous ratio, i.e. the factor that relates  $StO_2$  to  
24 arterial oxygen saturation  $SaO_2$  and to venous oxygen saturation  $SvO_2$  [41] All of this is really  
25 possible. Hence the physiologic interpretation of this finding is open.  
26  
27  
28  
29  
30  
31  
32  
33  
34  
35  
36  
37  
38

39  
40 In conclusion, changes in brain oxygenation and cerebral blood flow during transition were  
41 successfully and simultaneously measured by the BabyLux device in most infants. The dataset is  
42 sufficiently large to be robust. The  $StO_2$ , TOE and BFI values were plausible, although the decrease  
43 in  $CMRO_2I$  was unexpected and most likely indicate some oxygenation-level dependent error in one  
44 or more of the measured variables.  
45  
46  
47  
48  
49

50  
51 These preliminary results from the clinical application of the BabyLux device are encouraging in  
52 terms of reliability, safety, and feasibility.  
53  
54  
55  
56  
57

#### 58 **ACKNOWLEDGEMENTS:**

59  
60

1  
2  
3 The authors would like to thank all collaborators in the BabyLux consortium and the parents of infants  
4 included in the investigation. We are deeply thankful to Francesca Dessimone and Silvia Pisoni, who  
5  
6 helped us in recruiting the sample.  
7  
8

9  
10 **FUNDING:**

11  
12 This research was funded by the European Commission Competitiveness for Innovation Program  
13 (grant agreement no. 620996) as part of the project “An optical neuro-monitor of cerebral oxygen  
14 metabolism and blood flow for neonatology (BabyLux)”.

15  
16  
17 We also acknowledge funding from Fundació CELLEX Barcelona and “Severo Ochoa” Programme  
18 (SEV-2015-0522).  
19

20  
21  
22 **COMPETING INTERESTS:**

23  
24 Udo Weigel is the CEO, has equity ownership in HemoPhotonics S.L. and is an employee in the  
25 company. His role in the project has been defined by the project objectives, tasks and work-packages  
26 and was reviewed by the European Commission.  
27  
28  
29  
30  
31  
32  
33  
34  
35  
36  
37  
38  
39  
40  
41  
42  
43  
44  
45  
46  
47  
48  
49  
50  
51  
52  
53  
54  
55  
56  
57  
58  
59  
60

1  
2  
3  
4  
5  
6  
7  
8  
9  
10  
11  
12  
13  
14  
15  
16  
17  
18  
19  
20  
21  
22  
23  
24  
25  
26  
27  
28  
29  
30  
31  
32  
33  
34  
35  
36  
37  
38  
39  
40  
41  
42  
43  
44  
45  
46  
47  
48  
49  
50  
51  
52  
53  
54  
55  
56  
57  
58  
59  
60

### **What is already known on this topic:**

- A continuous and non-invasive monitoring of cerebral perfusion and oxygenation has been searched for better understanding the pathogenesis of brain damage in neonates.
- Time-resolved-near-infrared spectroscopy has the ability to resolve path-lengths of photons and to separate the absorption and scattering coefficients allowing for measurements of tissue oxygen saturation.
- Diffuse-correlation spectroscopy relies on the interaction between long coherence laser light and moving scatterers, allowing for quantification of microvascular blood flow and calculation of a blood flow index.

### **What this study adds:**

- The BabyLux device combines, for the first time, two photonic technologies (TRS/DCS), providing a non-invasive, cot-side and continuous monitoring of tissue oxygenation and blood flow.
- Cerebral tissue oxygenation and blood flow values obtained by BabyLux in term infants during transition after birth, were plausible.
- The cerebral metabolic rate of oxygen that was calculated from this data, decreased significantly: this was unexpected and the interpretation remains open.

**References:**

- 1 Joseph RM, O'Shea TM, Allred EN, et al. Neurocognitive and Academic Outcomes at Age 10 Years of Extremely Preterm Newborns. *Pediatrics* 2016;137(4).
- 2 Volpe JJ. Brain injury in premature infants: a complex amalgam of destructive and developmental disturbances. *Lancet Neurol* 2009;8(1):110-24.
- 3 Takami T, Sunohara D, Kondo A, et al. Changes in cerebral perfusion in extremely LBW infants during the first 72 h after birth. *Pediatr Res* 2010;68(5):435–9.
- 4 Deulofeut R, Critz A, Adams-Chapman I, Sola A. Avoiding hyperoxia in infants  $\leq 1250$  g is associated with improved short- and long-term outcomes. *Journal of Perinatology* 2006;26(11):700–5.
- 5 Greisen G, Leung T, Wolf M. Has the time come to use NIRS as a routine clinical tool in preterm infants undergoing intensive care? *Philos Trans A Math Phys Eng Sci.*2011;369(1955):4440-51.
- 6 Hyttel-Sorensen S, Greisen G, Als-Nielsen B, et al. Cerebral near-infrared spectroscopy monitoring for prevention of brain injury in very preterm infants. *Cochrane Database Syst Rev.* 2017;4;9:CD011506.
- 7 Hyttel-Sorensen S, Pellicer A, Alderliesten T, et al. Cerebral near infrared spectroscopy oximetry in extremely preterm infants: phase II randomised clinical trial. *BMJ* 2015;350:g7635.
- 8 Torricelli A, Contini D, Pifferi A, et al. Time domain functional NIRS imaging for human brain mapping, *Neuroimage* 2014;85Pt1:28-50.
- 9 Durduran T, Zhou C, Buckley EM, et al. Optical measurement of cerebral hemodynamics and oxygen metabolism in neonates with congenital heart defects. *J Biomed Opt* 2010;15(3):037004.

- 1  
2  
3 10 Durduran, T and Yodh AG, Diffuse correlation spectroscopy for non-invasive, micro-vascular  
4 cerebral blood flow measurement. *Neuroimage* 2014;85 Pt 1: p. 51-63.  
5  
6  
7  
8 11 Lin PY, Roche-Labarbe N, Dehaes M, et al. Non-invasive optical measurement of cerebral  
9 metabolism and hemodynamics in infants. *J Vis Exp* 2013;(73):e4379.  
10  
11  
12 12 Dehaes M, Aggarwal A, Lin PY, et al. Cerebral oxygen metabolism in neonatal hypoxic  
13 ischemic encephalopathy during and after therapeutic hypothermia. *J Cereb Blood Flow*  
14 *Metab* 2014;34(1):87-94.  
15  
16  
17  
18 13 Sybulski S, Maughan GB. Cortisol levels in umbilical cord plasma in relation to labor and  
19 delivery. *Am J Obstet Gynecol* 1976;125(2):236-8.  
20  
21  
22  
23 14 Irestedt L, Lagercrantz H, Hjerdahl P, et al. Fetal and maternal plasma catecholamine levels  
24 at elective cesarean section under general or epidural anesthesia versus vaginal delivery. *Am*  
25 *J Obstet Gynecol* 1982;142(8):1004-10.  
26  
27  
28  
29 15 Wang L, Zhang W, Zhao Y. The study of maternal and fetal plasma catecholamines levels  
30 during pregnancy and delivery. *J Perinat Med* 1999;27(3):195-8.  
31  
32  
33  
34 16 Perlman JM, Wyllie J, Kattwinkel J, et al. Part 7: Neonatal Resuscitation: 2015 International  
35 Consensus on Cardiopulmonary Resuscitation and Emergency Cardiovascular Care  
36 Science With Treatment Recommendations. *Circulation*. 2015;132(16 Suppl 1):S204-41.  
37  
38  
39  
40 17 Giovannella et al. The BabyLux device: baseline hemodynamic and optical properties of the  
41 newborn brain and the reproducibility of the results. *Optics InfoBase Conference Papers*  
42 2018; Part F90-OTS, 2p.  
43  
44  
45  
46 18 Cubeddu R, Pifferi A, Taroni P, et al. Experimental test of theoretical models for time-  
47 resolved reflectance, *Med Phys* 1996;23:1625  
48  
49  
50  
51 19 Prahl S, Optical Absorption of Hemoglobin, <http://omlc.ogi.edu/spectra/hemoglobin/>  
52  
53  
54  
55 20 Kreis R, Ernst T, Ross BD. Development of the human brain: in vivo quantification of  
56 metabolite and water content with proton magnetic resonance spectroscopy. *Magn Reson*  
57 *Med*. 1993;30(4):424-37.  
58  
59  
60


- 1  
2  
3 21 Durduran T, Choe R, Baker WB, Yodh AG. Diffuse Optics for Tissue Monitoring and  
4  
5 Tomography. *Rep Prog Phys*. 2010;73(7).  
6  
7  
8 22 Irwin D, Dong L, Shang Y, et al. Influences of tissue absorption and scattering on diffuse  
9  
10 correlation spectroscopy blood flow measurements. *Biomedical Optics Express* 2011; 2(7),  
11  
12 1969–1985.  
13  
14 23 Urlesberger B, Grossauer K, Pocivalnik M, et al. Regional oxygen saturation of the brain and  
15  
16 peripheral tissue during birth transition of term infants. *J Pediatr* 2010;157(5):740-4.  
17  
18 24 Baik N, Urlesberger B, Schwabegger B, et al. Reference Ranges for Cerebral Tissue Oxygen  
19  
20 Saturation Index in Term Neonates during Immediate Neonatal Transition after Birth.  
21  
22 *Neonatology* 2015;108(4):283-6.  
23  
24 25 Ziehenberger E, Urlesberger B, Binder-Heschl C, et al. Near-infrared spectroscopy  
26  
27 monitoring during immediate transition after birth: time to obtain cerebral tissue oxygenation.  
28  
29 *J Clin Monit Comput*. 2018;32(3):465-469.  
30  
31  
32 26 Fauchère JC, Schulz G, Haensse D, et al. Near-infrared spectroscopy measurements of  
33  
34 cerebral oxygenation in newborns during immediate postnatal adaptation. *J Pediatr* 2010  
35  
36 Mar;156(3):372-6.  
37  
38 27 Almaazmi M, Schmid MB, Havers S, et al. Cerebral near-infrared spectroscopy during  
39  
40 transition of healthy term newborns. *Neonatology* 2013;103(4):246–51.  
41  
42 28 Isobe K, Kusaka T, Fujikawa Y, et al. Measurement of cerebral oxygenation in neonates after  
43  
44 vaginal delivery and cesarean section using full-spectrum near-infrared spectroscopy. *Comp*  
45  
46 *Biochem Physiol A Mol Integr Physiol* 2002; 132: 133–138.  
47  
48 29 Urlesberger B, Kratky E, Rehak T, et al. Regional oxygen saturation of the brain during birth  
49  
50 transition of term infants: comparison between elective cesarean and vaginal deliveries. *J*  
51  
52 *Pediatr* 2011;159(3):404–8.  
53  
54  
55  
56  
57  
58  
59  
60

- 1  
2  
3 30 Pichler G, Binder C, Avian A, et al. Reference ranges for regional cerebral tissue oxygen  
4 saturation and fractional oxygen extraction in neonates during immediate transition after birth.  
5  
6  
7  
8  
9  
10 31 Dawson JA, Kamlin CO, Vento M, et al. Defining the reference range for oxygen saturation  
11 for infants after birth. *Pediatrics* 2010;125(6):e1340-7.  
12  
13  
14 32 Noori S, Wlodaver A, Gottipati V, et al. Transitional changes in cardiac and cerebral  
15 hemodynamics in term neonates at birth. *J Pediatr* 2012;160:943–948.  
16  
17  
18 33 Dullenkopf A, Frey B, Baenziger O, et al. Measurement of cerebral oxygenation state in  
19 anaesthetized children using the INVOS 5100 cerebral oximeter. *Paediatr Anaesth* 2003;  
20  
21  
22  
23  
24  
25 34 Morris N, Pichler G, Pocivalnik M, et al. Cerebral regional oxygen saturation (crSO<sub>2</sub>): are  
26 different sensors comparable? *Paediatr Anaesth* 2012;22(11):1132-4.  
27  
28  
29  
30 35 Montaldo P, De Leonibus C, Giordano L, et al. Cerebral, renal and mesenteric regional oxygen  
31 saturation of term infants during transition. *J Pediatr Surg* 2015;50(8):1273–7.  
32  
33  
34  
35 36 Dehaes M, Aggarwal A, Lin PY, et al. Cerebral oxygen metabolism in neonatal hypoxic  
36 ischemic encephalopathy during and after therapeutic hypothermia. *J Cereb Blood Flow*  
37  
38  
39  
40  
41  
42 37 Farzam P, Buckley E, Lin P, et al. Shedding light on the neonatal brain: probing cerebral  
43 hemodynamics by diffuse optical spectroscopic methods. *Scientific Reports* 2007;1–10.  
44  
45  
46 38 Giovannella M, Andresen B, Andersen JB, et al. Validation and potential calibration of diffuse  
47 correlation spectroscopy versus (15)O–water positron emission tomography on neonatal  
48 piglets, *fNIRS2018*, Society for functional near-infrared spectroscopy, Tokyo, 2018  
49  
50  
51  
52  
53 39 Andersen JB, Lindberg U, Olesen OV, et al. Hybrid PET/MRI imaging in healthy unsedated  
54 newborn infants with quantitative rCBF measurements using (15)O-water PET. *J Cereb Blood*  
55  
56  
57  
58  
59  
60

1  
2  
3 40 Greisen G. Cerebral blood flow in preterm infants during the first week of life. *Acta Paediatr*  
4  
5 *Scand.*1986;75(1):43-51.  
6


7  
8 41 Wong FY, Alexiou T, Samarasinghe T, et al. Cerebral arterial and venous contributions to  
9  
10 tissue oxygenation index measured using spatially resolved spectroscopy in newborn lambs.  
11  
12 *Anesthesiology* 2010;113(6):1385-91.  
13  
14  
15  
16  
17  
18  
19  
20  
21  
22  
23  
24  
25  
26  
27  
28  
29  
30  
31  
32  
33  
34  
35  
36  
37  
38  
39  
40  
41  
42  
43  
44  
45  
46  
47  
48  
49  
50  
51  
52  
53  
54  
55  
56  
57  
58  
59  
60

1  
2  
3  
4  
5  
6  
7  
8  
9  
10  
11  
12  
13  
14  
15  
16  
17  
18  
19  
20  
21  
22  
23  
24  
25  
26  
27  
28  
29  
30  
31  
32  
33  
34  
35  
36  
37  
38  
39  
40  
41  
42  
43  
44  
45  
46  
47  
48  
49  
50  
51  
52  
53  
54  
55  
56  
57  
58  
59  
60



Box-plot of SpO<sub>2</sub> changes over time.


142x80mm (300 x 300 DPI)



Box-plot of StO<sub>2</sub> changes over time.


142x80mm (300 x 300 DPI)

1  
2  
3  
4  
5  
6  
7  
8  
9  
10  
11  
12  
13  
14  
15  
16  
17  
18  
19  
20  
21  
22  
23  
24  
25  
26  
27  
28  
29  
30  
31  
32  
33  
34  
35  
36  
37  
38  
39  
40  
41  
42  
43  
44  
45  
46  
47  
48  
49  
50  
51  
52  
53  
54  
55  
56  
57  
58  
59  
60



Box-plot of BFI changes over time.


142x80mm (300 x 300 DPI)



Box-plot of TOE changes over time.

142x80mm (300 x 300 DPI)

1  
2  
3  
4  
5  
6  
7  
8  
9  
10  
11  
12  
13  
14  
15  
16  
17  
18  
19  
20  
21  
22  
23  
24  
25  
26  
27  
28  
29  
30  
31  
32  
33  
34  
35  
36  
37  
38  
39  
40  
41  
42  
43  
44  
45  
46  
47  
48  
49  
50  
51  
52  
53  
54  
55  
56  
57  
58  
59  
60



Box-plot of CMRO<sub>2</sub>I changes over time.

142x80mm (300 x 300 DPI)

**TITLE:**

Cerebral oxygenation and blood flow in term infants during postnatal transition: BabyLux project.


**SUPPLEMENTARY MATERIAL:**

Figure 1a: SpO<sub>2</sub> by subject

view Only




Figure 2a: StO<sub>2</sub> by subject




Figure 3a: BFI by subject




Figure 4a: TOE by subject




Figure 5a: CMRO<sub>2</sub>I by subject